



Magnetic Resonance Imaging (MRI) Compataibility Statement

The following SpineCraft's implant systems are fabricated from non-magnetic pure titanium, titanium alloys, tantalum and/or polymeric material (PEEK). They are not affected by the magnetic fields used in MRI with no heating, movement, or dislodgment. These implants are MR-CONDITIONAL using the terminology specified in the American Society for Testing and Materials (ASTM) International, Designation: F2503-05. Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment:

- APEX Spine System Implants
- AIM MIS SpineSystem Implants
- ORIO Interbody Fusion Cages
- ALTUM ACP System Implants

RE: SpineCraft MRI Compatyibility Document on Rev 2 - 5/14/2016

A handwritten signature in black ink, appearing to read 'Wagdy Asaad', is written over a horizontal line. There are some red ink marks above and below the signature.

Wagdy Asaad, MD
R&D Department Head
SpineCraft

June 16, 2016